

#### Site description

We have access to a list of dates when the diesel generator is in use at the military station, corresponding data are removed.

The railway is finished since may 2012 and there is no more car access to the station except for scientists and a for the military station.

#### Documentation

An online weekly check list of instrument operational sate is now filled in additionally to the paper list.

#### Inlets

Three total air inlets are operational (heated) since 2012 with identical designs and flow rates. The interstitial inlet are not used except during intensive campaigns for which the interstitial aérosol data are filtered from the files which are submitted to the data bases.

#### SMPS

No diffusion dryer is used in front of the SMPS since the measured RH at the entrance of the instrument does not exceed 40% all year long, due to differences between outside and inside temperatures.

#### OPC

The GRIMM OPC 1.108 was replaced with an OPC WELAS 2100 since January 2013